


CAEN n
Tools for Discovery

Register your device

Register your device to your **MyCAEN+** account and get access to our customer services, such as notification for new firmware or software upgrade, tracking service procedures or open a ticket for assistance. **MyCAEN+** accounts have a dedicated support service for their registered products. A set of basic information can be shared with the operator, speeding up the troubleshooting process and improving the efficiency of the support interactions.

MyCAEN+ dashboard is designed to offer you a direct access to all our after sales services. Registration is totally free, to create an account go to <https://www.caen.it/become-mycaenplus-user> and fill the registration form with your data.



<https://www.caen.it/become-mycaenplus-user/>

Technical Information Manual

Revision n. 5
26 July 2011

MOD. V1785
8 CH. DUAL RANGE
PEAK ADC
MANUAL REV.5

NPO:
00115/05:V1785.MUTx/05

CAEN will repair or replace any product within the guarantee period if the Guarantor declares that the product is defective due to workmanship or materials and has not been caused by mishandling, negligence on behalf of the User, accident or any abnormal conditions or operations.


CAEN declines all responsibility for damages or injuries caused by an improper use of the Modules due to negligence on behalf of the User. It is strongly recommended to read thoroughly the CAEN User's Manual before any kind of operation.



CAEN reserves the right to change partially or entirely the contents of this Manual at any time and without giving any notice.

Disposal of the Product

The product must never be dumped in the Municipal Waste. Please check your local regulations for disposal of electronics products.



MADE IN ITALY : We stress the fact that all the boards are made in Italy because in this globalized world, where getting the lowest possible price for products sometimes translates into poor pay and working conditions for the people who make them, at least you know that who made your board was reasonably paid and worked in a safe environment. (this obviously applies only to the boards marked "MADE IN ITALY", we can not attest to the manufacturing process of "third party" boards).

TABLE OF CONTENTS

1. GENERAL DESCRIPTION.....	8
1.1. OVERVIEW	8
1.2. BLOCK DIAGRAM.....	9
2. PRINCIPLES OF OPERATION.....	10
2.1. PEAK SECTIONS.....	10
2.2. ANALOG TO DIGITAL CONVERSION	12
2.3. ZERO SUPPRESSION	13
2.4. OVERFLOW SUPPRESSION	15
2.5. MULTIPLE EVENT BUFFER (MEB)	15
2.6. EVENT COUNTER.....	16
2.7. BUSY LOGIC	16
2.8. RESET LOGIC.....	17
2.9. FAST CLEAR	17
3. TECHNICAL SPECIFICATIONS.....	19
3.1. PACKAGING.....	19
3.2. POWER REQUIREMENTS	19
3.3. FRONT PANEL.....	20
3.4. EXTERNAL CONNECTORS.....	21
3.4.1. <i>INPUT</i> connectors	21
3.4.2. <i>CONTROL</i> connectors.....	21
3.4.3. <i>GATE /COMM</i> connectors	21
3.5. OTHER FRONT PANEL COMPONENTS	22
3.5.1. <i>Displays</i>	22
3.5.2. <i>Switches</i>	22
3.6. INTERNAL HARDWARE COMPONENTS	23
3.6.1. <i>Switches</i>	23
3.6.2. <i>Jumpers</i>	23
3.7. TECHNICAL SPECIFICATION TABLE	25
4. VME INTERFACE.....	26
4.1. ADDRESSING CAPABILITY.....	26
4.1.1. <i>Addressing via Base Address</i>	26
4.1.2. <i>Addressing examples</i>	27
4.1.3. <i>MCST/CBLT addressing</i>	27
4.1.4. <i>MCST/CBLT addressing examples</i>	29
4.2. INTERRUPTER CAPABILITY	30

4.2.1.	<i>Interrupt Status/ID</i>	30
4.2.2.	<i>Interrupt Level</i>	30
4.2.3.	<i>Interrupt Generation</i>	30
4.2.4.	<i>Interrupt Request Release</i>	30
4.3.	DATA TRANSFER CAPABILITY	30
4.4.	REGISTER ADDRESS MAP.....	31
4.5.	OUTPUT BUFFER REGISTER	34
4.5.1.	<i>Data storage</i>	36
4.5.2.	<i>Data readout</i>	36
4.6.	FIRMWARE REVISION REGISTER	37
4.7.	GEO ADDRESS REGISTER.....	38
4.8.	MCST/CBLT ADDRESS REGISTER.....	39
4.9.	BIT SET 1 REGISTER	39
4.10.	BIT CLEAR 1 REGISTER.....	40
4.11.	INTERRUPT LEVEL REGISTER.....	40
4.12.	INTERRUPT VECTOR REGISTER	41
4.13.	STATUS REGISTER 1	41
4.14.	CONTROL REGISTER 1	43
4.15.	ADDRESS DECODER HIGH REGISTER	44
4.16.	ADDRESS DECODER LOW REGISTER	44
4.17.	SINGLE SHOT RESET REGISTER	44
4.18.	MCST/CBLT CONTROL REGISTER	44
4.19.	EVENT TRIGGER REGISTER.....	45
4.20.	STATUS REGISTER 2	46
4.21.	EVENT COUNTER_LOW REGISTER.....	46
4.22.	EVENT COUNTER_HIGH REGISTER	47
4.23.	INCREMENT EVENT REGISTER	47
4.24.	INCREMENT OFFSET REGISTER	48
4.25.	FAST CLEAR WINDOW REGISTER	48
4.26.	BIT SET 2 REGISTER	49
4.27.	BIT CLEAR 2 REGISTER.....	51
4.28.	W MEMORY TEST ADDRESS REGISTER	51
4.29.	MEMORY TEST WORD_HIGH REGISTER	51
4.30.	MEMORY TEST WORD_LOW REGISTER	52
4.31.	CRATE SELECT REGISTER	52
4.32.	TEST EVENT WRITE REGISTER	52
4.33.	EVENT COUNTER RESET REGISTER	53
4.34.	R MEMORY TEST ADDRESS REGISTER.....	53

4.35.	SW COMM REGISTER	53
4.36.	SLIDE CONSTANT REGISTER	54
4.37.	AAD REGISTER.....	54
4.38.	BAD REGISTER	54
4.39.	THRESHOLDS MEMORY	54
4.40.	ROM MEMORY	55
5.	OPERATING MODES	56
5.1.	SAFETY INFORMATION	56
5.2.	PRODUCT DAMAGE PRECAUTIONS	56
5.3.	INSTALLATION.....	56
5.4.	POWER ON SEQUENCE	56
5.5.	POWER ON STATUS	57
5.6.	OPERATION SEQUENCE.....	57
5.7.	TEST MODES	58
5.7.1.	<i>Random Memory Access Test Mode</i>	58
5.7.2.	<i>Acquisition Test Mode</i>	59
5.8.	BLOCK TRANSFER MODE	59
5.9.	ADVANCED SETTING AND READOUT MODES	60
5.9.1.	<i>Chained Block Transfer Mode</i>	61
5.9.2.	<i>Multicast Commands</i>	62
6.	REFERENCES	63
APPENDIX A	64
	<i>VME INTERFACE TIMING</i>	64
A.1	VME CYCLE TIMING IN D16/D32 MODE	65
A.2	VME CYCLE TIMING IN BLT / CBLT MODE.....	65
A.3	VME CYCLE TIMING IN MBLT / CBLT64 MODE	66

LIST OF FIGURES

FIG. 1.1: MODEL V1785 BLOCK DIAGRAM	9
FIG. 2.1: SIMPLIFIED BLOCK DIAGRAM OF THE PEAK SECTION	10
FIG. 2.2: SIGNAL CONVERSION TIMING	11
FIG. 2.3: SIGNAL CONVERSION EXAMPLE	12
FIG. 2.4: BLOCK DIAGRAM OF THE SLIDING SCALE SECTION	13
FIG. 2.5: ZERO SUPPRESSION (BIT 8 OF BIT SET 2 REGISTER = 0, DEFAULT SETTING).....	14

FIG. 2.6: ZERO SUPPRESSION (BIT 8 OF BIT SET 2 REGISTER =1)	14
FIG. 2.7: MULTI-EVENT BUFFER: WRITE POINTER AND READ POINTER.....	15
FIG. 2.8: FAST CLEAR WINDOW	18
FIG. 3.1: MODEL V1785 FRONT PANEL	20
FIG. 3.2: V1785 COMPONENT LOCATION (COMPONENT SIDE)	24
FIG. 4.3: BASE ADDRESSING: EXAMPLE 1.....	27
FIG. 4.4: MCST/CBLT ADDRESSING EXAMPLE	29
FIG. 4.5: OUTPUT BUFFER: THE HEADER.....	34
FIG. 4.6: OUTPUT BUFFER: THE DATA WORD FORMAT	34
FIG. 4.7: OUTPUT BUFFER: THE END OF BLOCK.....	34
FIG. 4.8: OUTPUT BUFFER: NOT VALID DATUM	35
FIG. 4.9: DATA READOUT AND SUPPRESSION	36
FIG. 4.10: MULTI-EVENT BUFFER: DATA STRUCTURE EXAMPLE.....	37
FIG. 4.11: FIRMWARE REVISION REGISTER.....	37
FIG. 4.12: GEOGRAPHICAL ADDRESS REGISTER	38
FIG. 4.13: MCST/CBLT ADDRESS REGISTER.....	39
FIG. 4.14: BIT SET 1 REGISTER	39
FIG. 4.15: INTERRUPT LEVEL REGISTER	40
FIG. 4.16: INTERRUPT VECTOR REGISTER.....	41
FIG. 4.17: STATUS REGISTER	41
FIG. 4.18: CONTROL REGISTER 1	43
FIG. 4.19: ADER HIGH REGISTER	44
FIG. 4.20: ADER LOW REGISTER.....	44
FIG. 4.21: MCST ADDRESS REGISTER.....	44
FIG. 4.22: EVENT TRIGGER REGISTER	45
FIG. 4.23: STATUS REGISTER 2	46
FIG. 4.24: EVENT COUNTER LOW REGISTER.....	47
FIG. 4.25: EVENT COUNTER HIGH REGISTER.....	47
FIG. 4.26: FAST CLEAR WINDOW REGISTER	48
FIG. 4.27: BIT SET 2 REGISTER	49
FIG. 4.28: W MEMORY TEST ADDRESS REGISTER	51
FIG. 4.29: TEST WORD_HIGH REGISTER.....	51
FIG. 4.30: TEST WORD_LOW REGISTER	52
FIG. 4.31: CRATE SELECT REGISTER.....	52
FIG. 4.32: TEST EVENT WRITE REGISTER	53
FIG. 4.33: R MEMORY TEST ADDRESS REGISTER	53

FIG. 4.34: SLIDE CONSTANT REGISTER	54
FIG. 4.35: AAD REGISTER	54
FIG. 4.36: BAD REGISTER.....	54
FIG. 4.37: THRESHOLD REGISTER.....	55
FIG.A.1: VME CYCLE TIMING IN D16 MODE	65
FIG.A.2: VME CYCLE TIMING IN BLT/CBLT MODE	65
FIG.A.3: VME CYCLE TIMING IN MBLT/CBLT64 MODE	66

LIST OF TABLES

TABLE 3.1: MOD. V1785 POWER REQUIREMENTS	19
TABLE 3.2: MODEL V1785 MAIN TECHNICAL SPECIFICATIONS.....	25
TABLE 4.1: MODULE RECOGNISED ADDRESS MODIFIER.....	26
TABLE 4.2: ADDRESS MAP FOR THE MODEL V1785.....	32
TABLE 4.3: ADDRESS MAP IN CBLT OPERATION	33
TABLE 4.4: ADDRESS MAP IN MCST OPERATIONS.....	33
TABLE 4.5: ROM ADDRESS MAP FOR THE MODEL V1785	55

1. General description

1.1. Overview

The Mod. V1785 is a 1-unit wide VME 6U module housing 8 Peak Sensing Analog-to-Digital Conversion channels. Each channel is able to detect and convert the peak value of the positive analog signals (with >50 ns risetime) fed to the relevant connectors. Input voltage range is $0 \div 4$ V. Each channel is processed by a x1 gain stage and by a x8¹ gain stage in parallel, then by the ADC stage: a dual input range is then featured: $0 \div 4$ V (1 mV LSB) and $0 \div 500$ mV (125 μ V LSB); this allows to avoid saturation with big input signals while increasing resolution with small ones.

The ADCs use a sliding scale technique in order to reduce the differential non-linearity. Programmable zero suppression, multievent buffer memory, trigger counter and test features complete the flexibility of the unit.

The module works in A24/A32 mode. The data transfer occurs in D16, D32, BLT32 or MBLT64 mode. The unit supports also the Chained Block Transfer (CBLT32/CBLT64) and the Multicast commands.

The VME interface is VME64X standard compliant and features the A24/A32 and MultiCast addressing modes. The data readout occurs either in D32, BLT32, MBLT64 mode, or in daisy chain with 32/64 bit Chained Block Transfers. The module features a fully programmable RORA interrupter.

The board is provided with the P1 and P2 VME connectors and fits into both V430 and standard VME 6U crates. It also supports the “live insertion”, allowing the User to insert (or remove) the board into (or from) the crate without switching it off.

¹ Not calibrated

1.2. Block diagram




Fig. 1.1: Model V1785 Block Diagram

2. Principles of operation

The board has 8 input channels and one GATE input common to all channels.

The peak values, received from the channel inputs when the GATE input signal is active, are converted into voltage levels by the Peak Sensing sections (PEAK sections) and then multiplexed and converted by two fast 12-bit ADC modules.

Each input signal is converted into two words, one featuring $125 \mu\text{V}$ LSB (500 mV full scale) and the other 1 mV LSB (4 V full scale).

Only the values that are above a programmable threshold (see § 2.3), do not cause overflow (see § 2.3) and are not killed (see § 2.3) will be stored in a dual port data memory accessible via VME.

In the following functional sections and operation principles of the module are described in some detail. The block diagram of the module can be found in Fig. 1.1.

2.1. PEAK sections

The module hosts 16 PEAK sections: a simplified block diagram of a PEAK section is reported in Fig. 2.1. The peak detection is based on the charging of a capacitor at constant current.

The GATE signal closes the switch SW1 thus allowing the capacitor C1 to be charged as the diode D1 is forward-biased by the signal. The signal is amplified and fed to the multiplexer. As the SW1 is open again, the signal is digitised by the 12-bit ADCs. After digitisation the SW2 switch is closed by the CLEAR signal which allows the discharge of the capacitor C1. Both the GATE and CLEAR signals are controlled by the CONTROL LOGIC section.




Fig. 2.1: Simplified block diagram of the PEAK section




Fig. 2.2: Signal conversion timing

The signal conversion timing is shown in Fig. 2.2. The diagram includes four different time ranges: idle, data acquisition, settling time, digitisation and clear.

While the conversion logic is **idle**, the occurrence of a GATE pulse starts the **acquiring data** phase, during which the PEAK output increases according to the input signal until the first peak is reached. As the peak is reached, the peak value is held by means of the capacitor C1 until the end of the digital conversion (**digitisation**) which starts about 600 ns (**settling time**) after the end of the GATE signal and takes about 6 μ s. After the digital conversion, the **clear** phase takes place by a fast capacitor discharge (about 600 ns) which makes the conversion logic idle again.




Fig. 2.3: Signal conversion example

2.2. Analog to digital conversion

The output of each PEAK section is multiplexed, by group of 4 channels, and subsequently converted by two fast 12 bit ADCs, each of which operates the conversion on a group of 8 channels (Block A and Block B ADCs). The ADC section supports the sliding scale technique to reduce the differential non-linearity (see references [1], [2]). This technique (see Fig. 2.4) consists in adding a known value to the analog level to be converted, thus spanning different ADC conversion regions with the same analog value. The known level is then digitally subtracted after the conversion and the final value is sent to the threshold comparator.

If the sliding scale is enabled, it reduces slightly the dynamic range of the ADC: the 12-bit digital output is valid from 0 to 3840, while the values from 3841 to 4095 are not correct.




Fig. 2.4: Block diagram of the sliding scale section

2.3. Zero suppression

The output of the ADC is fed to a threshold comparator to perform the zero suppression. If the converted value from a channel is greater than (or equal to) the relevant low threshold value set via VME in the Thresholds memory (Base Address + 0x1080 ÷ 0x10BF, see § 4.39), the result is fed to the dual port memory and will be available for the readout.

If the converted value is lower than the threshold, the value is stored in the memory only if the LOW TRESHOLD PROG. bit of the Bit Set 2 Register is set to 1 (see § 4.26). The fact that the converted value was under the threshold is also flagged in the datum stored in the memory, where the bit 13 (UNDERTHRESHOLD) of the 16-bit data word is set to 1. The Thresholds memory allows to set a low threshold value for each channel. Default setting corresponds to thresholds not defined. By setting the bit 8 in the Bit Set 2 Register it is possible to program the Threshold values in 16 ADC counts steps over the entire full scale range or in 2 ADC counts steps over 1/8 of full scale range*. In more detail, if Bit 8 = 0 (default value) the comparison is performed between the 8 MSB of each 12 bit converted value and the 8 bit threshold value which is stored in the relevant register as illustrated in Fig. 2.5. The threshold values can be programmed over the entire full scale range.

* This feature is available from firmware releases 5.1; for earlier firmware releases the thresholds can be programmed only in 16 ADC counts steps (as illustrated in Fig. 2.4).




Fig. 2.5: Zero suppression (Bit 8 of Bit Set 2 Register = 0, default setting)

if Bit 8 = 1 (in the Bit Set 2 Register) the comparison is performed between the bit 1...8 of each 12 bit converted value and the 8 bit threshold value which is stored in the relevant register as illustrated in the figure below (converted value is under threshold if the value written in the 1...8 bits is smaller than the threshold value and 9...11 bits are 0). The threshold values can be programmed over 1/8 of full scale range.




Fig. 2.6: Zero suppression (Bit 8 of Bit Set 2 Register =1)

The comparison is resumed in the following table:

STEP_TH Bit (Bit 8 of Bit Set 2 Register)	Comparison
1	ADC CONVERTED VALUE < THRESHOLD VALUE x 2
0	ADC CONVERTED VALUE < THRESHOLD VALUE x 16

If the result of the comparison is true and the Bit 4 (LOW THRESHOLD PROG) of the Bit Set 2 Register is set to 0, data are skipped. If the Bit 4 of the Bit Set 2 Register is set to 1, the true result of the comparison is signalled by Bit 13 (UNDERTHRESHOLD) = 1 in the loaded data 16 bit word.

The content of the Threshold Register includes also a KILL bit, which allows to abort the memorisation of the datum even if it is higher than the threshold set in the register. This bit can thus be used to disable some channels. Refer to § 4.39 for further details.

The threshold values are lost only after switching the board off (a reset operation does not affect the threshold values).

2.4. Overflow suppression

The overflow suppression allows to abort the memorisation of data which originated an ADC overflow. The control logic provides to check if the output of the ADC is in overflow and, in the case, the value is not stored in the memory.

The overflow suppression can be disabled by means of the OVER RANGE PROG bit of the Bit Set 2 Register (see § 4.26): if this bit is set to 1, all the data, independently from the fact that they caused ADC overflow or not, are stored in the memory. In this case, the 16-bit word stored in the memory will have the bit 12 (OVERFLOW) set to 1 (see § 4.5).

2.5. Multiple Event Buffer (MEB)

After the conversion, if there is at least one converted value above the programmed threshold, not causing overflow and not killed, the control logic stores it in the Multi-Event Buffer (MEB).

The Multi-Event Buffer is a Dual Port Memory (34 Words/event) which can store up to 32 events. It is mapped at the VME address: Base Address + 0x0000÷0x07FC (see also § 4.5).

In order to trace the event flow, two pointers (Read and Write pointer) are employed.

The Read Pointer points to the active read buffer. The Write pointer is incremented automatically via hardware at the end of the channels conversion, while the Read pointer can be either incremented automatically (AUTO INCR. bit of the Bit Set 2 Register set to 1; see § 4.26) or via write access to one of two dummy registers, Increment Event and Increment Offset Registers (see § 4.23 and 4.24). These allow to move the readout pointer to the next event in the output buffer or to the next word, respectively.




Fig. 2.7: Multi-Event Buffer: Write pointer and Read pointer

The MEB can be either in a "Full", a "Not empty" or an "Empty" status.

When the 5MSB of the Read pointer and the 5MSB of the Write pointer are different (i.e. point to different events), the MEB is in a "Not empty" status.

When the Read pointer and the Write pointer are equal, the MEB can be either in a "Full" or an "Empty" status. The MEB is full or empty according to the last increment pointer operation performed: if the last increment is the one of the Write pointer, the MEB is Full; if the last increment is the one of the Read pointer, the MEB is Empty.

The status of the MEB is monitored via two Registers, the Status Register 1 and the Status Register 2 (see § 4.13 and § 4.20, respectively).

After the conversion, the accepted data (i.e. the converted values above the programmed threshold, not causing overflow and not killed) are stored in the active event buffer (i.e. the one pointed by the write pointer) in subsequent 32-bit words. These are organised in events. Each event consists of a Header (see Fig. 4.5), a block of data words (Fig. 4.6) and an End-Of-Block (EOB) word (Fig. 4.7).

Each event contains thus from a minimum of 3 32-bit words (Header, one data word and EOB) to a maximum of 34 32-bit words (Header, 32 data words and EOB).

Even if there are no accepted data, the User can choose to store in the MEB the Header and the EOB relative to the event (see EMPTY PROG bit of the Bit Set 2 Register, see § 4.26): in this case the event is constituted by 2 32-bit words only.

2.6. Event Counter

The module houses a 24-bit counter that counts the number of GATE signals that the module has received.

The Event Counter can work in two different modes, which can be selected via the Bit 14 (ALL TRG) of the Bit Set 2 Register (see § 4.26):

Mode A (ALL TRG = 1): it counts all events (default);
Mode B (ALL TRG = 0): it counts only the accepted events.

In the first case (Mode A), the Event Counter is increased each time a pulse is sent through the GATE input.

In the second case (Mode B), the Event Counter is increased each time a pulse, sent through the GATE input, is accepted (i.e. VETO, FCLR and BUSY are not active).

The value of the Event Counter is stored in the EOB of the Multi-Event Buffer (see § 4.5). The Event Counter is also stored in two registers, the Event Counter_Low and Event Counter_High Registers, which respectively contain the 16LSBs and the 8MSBs of the Event Counter (see § 4.21 and § 4.22).

2.7. Busy Logic

The board is BUSY either during the conversion sequence or during the reset of the analog section or when the MEB is not ready to accept data (MEB Full) or when the board is in Random Memory Access Test mode (see § 5.7.1).

On the occurrence of one of these conditions the front panel BUSY signal (CONTROL bus) is active, the red BUSY LED is on and the bit 2 (BUSY) and bit 3 (GLOBAL BUSY) of the Status Register 1 are set to 1 (see § 4.13).

The BUSY LED also lights up while the board is configuring (power ON).

Actually, each module sets to 1 its BUSY output after the leading edge of a pulse on the GATE input and releases it to 0 at the end of the conversion sequence. When the module is busy, it does not accept another GATE pulse.

The jumper J12 placed on the PCB (see Fig. 3.2) allows to select board behaviour in response to a BUSY status: if this jumper is set to EXTBSY, the acquisition is stopped as soon as any of the boards on the Control bus is BUSY; if the jumper is set to INTBSY, acquisition is stopped as the board is BUSY.

2.8. Reset Logic

Three different types of RESET operations can be distinguished, according to the effects they have on the module and particularly on the registers. They are:

- *Type A:* Data RESET
- *Type B:* Software RESET
- *Type C:* Hardware RESET

The **Data RESET** clears the data in the output buffer, resets the read and write pointers, the event counter and the peak sections. It does not affect the registers.

This type of RESET can be forwarded in two ways:

1. setting the Bit 2 (CLEAR DATA) of the Bit Set 2 Register to 1 (see § 4.26); the Reset is released via the Bit Clear 2 Register (see § 4.27);
2. sending a RESET pulse from the front panel with the Bit 4 (PROG RESET) of the Control Register 1 set to 0 (see § 4.14).

The **Software RESET** performs the same actions as the data RESET and, moreover, it resets the registers marked in the column SR (Software Reset) in Table 4.2.

This type of RESET can be forwarded in three ways:

1. setting the Bit 7 (SOFTWARE RESET) of the Bit Set 1 Register to 1 (see § 4.9): this sets the module to a permanent RESET status which is released only via write access, with the relevant bit set to 1, to the Bit Clear Register;
2. sending a RESET pulse from the front panel with the Bit 4 (PROG RESET) of the Control Register 1 set to 1 (see § 4.14);
3. performing a write access to the Single Shot Reset Register (see § 4.17): the RESET lasts as long as the write access itself.

The **Hardware RESET** performs the same actions as the Software RESET and, moreover, it resets further registers. All the registers reset by a Hardware RESET are marked in the column HR (Hardware Reset) in Table 4.2.

This type of RESET is performed:

1. at Power ON of the module;
2. via a VME RESET (SYS_RES).

At power ON or after a reset the module must thus be initialised.

2.9. FAST CLEAR

The FAST CLEAR of the module can be performed via the relevant front panel signal (see § 3.4.2). A FAST CLEAR signal, generated at any time within the FAST CLEAR window, i.e. between the leading edge of the GATE signal and the end of the programmable time value set in the Fast Clear Window Register (see § 4.25), aborts the conversion. Its minimum width must be 30 ns.

N.B.: since a FAST CLEAR operation implies a CLEAR CONVERSION cycle, a new GATE signal is accepted only if it occurs at least 600 ns after the leading-edge of the FAST CLEAR signal.




Fig. 2.8: Fast Clear window

3. Technical specifications

3.1. Packaging

The Model V1785 is housed in a 6U-high, 1U-wide VME unit. The board hosts the VME P1 and P2 connectors.

3.2. Power requirements

The power requirements of the V1785 module are as follows:

Table 3.1: Mod. V1785 power requirements

Power supply	V1785
+12 V	670 mA
-12 V	630 mA
+5 V	1.62 A

3.3. Front Panel




Fig. 3.1: Model V1785 front panel

3.4. External connectors

The location of the connectors is shown in Fig. 3.1. Their function and electro-mechanical specifications are listed in the following subsections.

3.4.1. INPUT connectors

Mechanical specifications:

8 LEMO 00 connectors.

Electrical specifications:

positive signals on 1 K Ω impedance; min. rise time: 50 ns. Input voltage range: 0 \div 4 V.

BLOCK A INPUT: input signals from channel 0 through channel 3.

BLOCK B INPUT: input signals from channel 4 through channel 7.

3.4.2. CONTROL connectors

Mechanical specifications:

four LEMO 00 type connectors.

FCLR: *Electrical specifications:* NIM std. input signal, 50 Ω impedance; min. FWHM: 30 ns.

Function: FAST CLEAR signal, accepted if sent within the so-called FAST CLEAR window; it clears the PEAK sections of the unit and aborts completely the conversion in progress.

RST: *Electrical specifications:* NIM std. input signal, 50 Ω impedance; min. FWHM: 30 ns.

Function: clears the PEAK sections, resets the Multi-Event Buffer status, stops pending ADCs conversions and, depending on the User's settings (see PROG RESET, § 4.14), may clear the control registers.

VETO: *Electrical specifications:* NIM std. input signal, 50 Ω impedance.

Function: inhibits the conversion of the PEAK signals.

BUSY: *Electrical specifications:* NIM output signal.
Function: indicates that the board is either converting or resetting or in MEMORY TEST mode or the MEB is full, BUSY status is also flagged by the bit 2 of the Status Register 1; when several boards are daisy-chained, the wired OR and wired NAND of BUSY signals can be read respectively on the BUSY+ and BUSY- lines of the CONTROL bus. The status of the BUSY+ bidirectional line is flagged by the bit 3 of the Status Register 1

3.4.3. GATE /COMM connectors

Mechanical specifications:

two bridged 00-type LEMO connectors.

Electrical specifications:

NIM std. input signals; high impedance; if this input is used a 50Ω termination is required; in daisy-chain configuration, the termination must be inserted on the last board of the chain.

GATE/COMM: *Function:* input signal, common to all channels, acting as the temporal window within which the peaks are detected.

3.5. Other front panel components

3.5.1. Displays

The front panel (refer to Fig. 3.1) hosts the following LEDs:

DTACK:	<i>Colour:</i> green. <i>Function:</i> DATA ACKNOWLEDGE command; it lights up each time a VME access is performed.
BUSY:	<i>Colour:</i> red. <i>Function:</i> it lights up each time the module is performing a conversion or resetting the analog section or in memory TEST mode or when the Multi-Event Buffer is full; it also lights up for a while at power ON to indicate that the board is configuring.
DRDY:	<i>Colour:</i> yellow. <i>Function:</i> it lights up when at least one event is present in the output buffer; it also lights up for a while at power ON to indicate that the board is configuring.
OVC/PWR:	<i>Colour:</i> green/orange. <i>Function:</i> it lights up green when the board is inserted into the crate and the crate is powered up; when it is orange, it indicates that there is an over-current status: in this case, remove the overload source, switch the module off and then switch it on again.

3.5.2. Switches

PWR:	<i>Type:</i> miniature flush plunger push-button switch. <i>Function:</i> after the insertion of the board into the crate, it allows to turn the board ON/OFF by pushing it with a pin. Refer to § 5.4 for the power ON procedure
-------------	--

3.6. Internal hardware components

The V1785 module is constituted by a motherboard with a piggy-back board plugged into it (see also Fig. 1.1 where the functional blocks hosted on the piggy-back board are pointed out). In the following some hardware setting components, located on the boards, are listed. Refer to Fig. 3.2 for their exact location on the PCB and their settings.

3.6.1. Switches

ROTARY SWITCHES:

Type: 4 rotary switches.

Function: they allow to select the VME address of the module. Please refer to Fig. 3.2 for their settings.

3.6.2. Jumpers

J12:

Function: it allows to select board behaviour in response to a BUSY status:

Position A (high): data acquisition is stopped as soon as any of the boards on the CONTROL Bus is BUSY;

Position B (low): data acquisition is stopped as the board is BUSY, independently from the status of the other boards on the CONTROL Bus.

Refer to Fig. 3.2 for the exact location of the jumper on the PCB and its setting.




Fig. 3.2: V1785 Component Location (component side)

3.7. Technical specification table

Table 3.2: Model V1785 main technical specifications

Inputs	8 channels, 1 kΩ impedance, positive polarity, DC coupling
Input range²	Dual range: 0 ÷ 4 V / 0 ÷ 500 mV
Resolution	12 bit ³
Gain	High range: 1 mV/count; Low range: 125 µV/count
Min. input voltage	1 mV
Reflections	<5% with 2 ns fall time input signals
Integral non linearity	± 0.1%
Differential non linearity	± 1.5%
Input offset	±2 mV
RMS Noise (typical)	0.7 counts (high range), 1.5 counts (low range)
Interchannel gain uniformity	± 4%
Interchannel Isolation	> 60 dB
Fast clear time	600 ns
Conversion time	2.8 µs / 8 Channel
Dead time	3.5 µs / 8 Channel
Zero suppression	Thresholds values programmable in: 16 ADC counts step over the entire FSR 2 ADC counts steps over 1/8 of FSR
Integral non linearity	± 0.1% of Full Scale Range ⁴ (T.B.C.)
GATE/COM input	Temporal window for peak detection
Control inputs	RST : resets PEAK sections, MEB status and control registers VETO : inhibits the conversion of the peaks FCLR : FAST CLEAR of PEAK sections and conversion
Control outputs	BUSY : indicates the presence of data
MultiEvent Buffer	32 events
Packaging	6U-high, 1U-wide VME unit
Addressing	A24 / A32 / Geographical / MultiCast
Register Access	D16 / D32 - Read / Write
Data Readout modes	D32 / BLT32 / MBLT64 / CBLT32 / CBLT64
Data throughput	Max: 40 Mbyte/s
Interrupt	7 level RORA Interrupter
Live Insertion	Yes
JAUX	Not used
Power Supplies	+5V, +12V, -12V
Compliance	VME / VME32 / VME64 / VME64X / VME64XP

² If Sliding Scale is used FSR is reduced from 4095 to 3840 counts

³ This value leads to a 15 bit dynamics

⁴ from 5% to 95% of FSR=3840 counts

4. VME interface

4.1. Addressing capability

The modules can be addressed in three different ways, specifically:

1. via Base Address;
2. via GEOgraphical address (to be used only for CBLT cycles);
3. via Multicast/Chained Block Transfer addressing mode.

4.1.1. Addressing via Base Address

The module works in A24/A32 mode. This implies that the module's address must be specified in a word of 24 or 32 bit. The Address Modifier codes recognised by the module are summarised in Table 4.1.

Table 4.1: Module recognised Address Modifier

A.M.	Description
0x3F	A24 supervisory block transfer (BLT)
0x3D	A24 supervisory data access
0x3C	A24 supervisory 64 bit block transfer (MBLT)
0x3B	A24 non privileged block transfer (BLT)
0x39	A24 non privileged User data access
0x38	A24 non privileged 64 bit block transfer (MBLT)
0x2F	Configuration Rom/Control & Status Register (CR/CSR)
0x0F	A32 supervisory block transfer (BLT)
0x0D	A32 supervisory data access
0x0C	A32 supervisory 64 bit block transfer (MBLT)
0x0B	A32 non privileged block transfer (BLT)
0x09	A32 non privileged data access
0x08	A32 non privileged 64 bit block transfer (MBLT)

The Base Address can be selected in the range:

0x00000000 \leftrightarrow 0xFF0000 A24 mode
 0x0000000000 \leftrightarrow 0xFFFF0000 A32 mode

The Base Address of the module can be fixed in two ways:

- by four rotary switches;
- by writing the Base Address in the ADER_HIGH and ADER_LOW registers.

The 4 rotary switches for Base Address selection are housed on two piggy-back boards plugged into the main printed circuit board (see Fig. 3.2).

To use this addressing mode the bit 4 of the Bit Set 1 Register (see § 4.9) must be set to 0. This is also the default setting.

The module Base Address can also be fixed by using the Ader_High and Ader_Low Registers. These two registers set respectively the A[31:24] and the A[23:16] VME address bits (see § 4.15 and 4.16).

To use this addressing mode bit 4 of the Bit Set 1 Register (see § 4.9) must be set to 1.

4.1.2. Addressing examples

The following is an example of Base Addressing for two V1785 boards inserted in a VME crate.




Fig. 4.3: Base Addressing: Example 1

If the board 1 and board 2 are respectively inserted in the slots 5 and 8 with the rotary switches for VME Base Addressing set as shown in the figure, the complete address of the registers of the two boards will be as follows:

Board 1:

Base addressing A32: 0xEE000000 + offset
Base addressing A24: 0x000000 + offset

Board 2:

Base addressing A32: 0xCC110000 + offset
Base addressing A24: 0x110000 + offset

4.1.3. MCST/CBLT addressing

When the Multicast/Chained Block Transfer addressing mode is adopted, the module works in A32 mode only. The Address Modifiers codes recognised by the module are:

AM=0x0F: A32 supervisory block transfer (CBLT)
AM=0x0D: A32 supervisory data access (MCST)
AM=0x0B: A32 User block transfer (CBLT)
AM=0x09: A32 User data access (MCST)

The boards can be accessed in Multicast Commands mode (MCST mode, see [4]), that allows to write in the registers of several boards at the same time by accessing the MCST Base Address in A32 only once.

The boards can be accessed in Chained Block Transfer mode (CBLT mode, see [4]) that allows to readout sequentially a certain number of contiguous boards in a VME crate. This access is allowed in BLT32 and BLT64 modes only to the MCST Base Address.

N.B.: The Base Address used for MCST and CBLT operations is the same, i.e. throughout this User's Manual the "MCST Base Address" identifies the same Address, used both for MCST commands (in Write only) and the CBLT Readout (in Read only, for the Output Buffer only).

The MCST Base Address must be set in a different way from the ordinary Base Address. Its most significant byte (i.e. bits 31 through 24) must be written in the MCST/CBLT Address Register (see § 4.8) and must be set in common to all boards belonging to the MCST/CBLT chain (i.e. all boards must have the same setting of the MCST/CBLT Base Address on bits 31 through 24). The default setting is 0xAA.

In CBLT and MCST operations, the IACKIN/IACKOUT daisy chain is used to pass a token from a board to the following one. The board which has received the token stores/sends the data from/to the master via CBLT/ MCST access. No empty slots must thus be left between the boards or, in alternative, empty slots can be left only in case VME crates with automatic IACKIN/IACKOUT short-circuiting are used.

Once the addresses have been set, the first and last board in a chain must have, respectively, only the FIRST_BOARD (F_B) and only the LAST_BOARD (L_B) bit set to 1 in the MCST Control Register (see § 4.18). On the contrary, all intermediate boards must have both the FIRST_BOARD and the LAST_BOARD bits set to 1 (active, intermediate) or both the FIRST_BOARD and the LAST_BOARD bits set to 0 (inactive). By default these bits are set to 0 (the board is inactive).

Board status	Board position in the chain	F_B bit	L_B bit
inactive	-	0	0
active	last	0	1
active	first	1	0
active	intermediate	1	1

Please note that in a chain there must be one (and only one) *first board* (i.e. a board with F_B bit set to 1 and the L_B bit set to 0) and one (and only one) *last board* (i.e. a board with F_B bit set to 0 and the L_B bit set to 1).

The complete address in A32 mode is:

A [31:24]	MCST/CBLT Address
A [23:16]	00
A [15:0]	offset

In MCST/CBLT operation it is possible to define more chains in the same crate, but each chain must have an address different from the other.

N.B.: In CBLT operation the data coming from different boards are tagged with the HEADER and with the EOB words containing the GEO address in the 5 MSB (see § 4.5). Since the module has not the PAUX connector, it is up to the User to write

the GEO address in the GEO register before executing the CBLT operation. If the GEO address is not written in the relevant register before performing the CBLT operation, it will not be possible to identify the module which the data are coming from.

4.1.4. MCST/CBLT addressing examples

The following is an example of MCST and CBLT addressing for four V1785 boards plugged into a VME crate. The steps to be performed to access the boards are as follows:

1. Set the MCST address (see § 4.8) for all boards via VME Base Address;
2. Set the bits F_B and L_B of the MCST Control Register (see § 4.18) according to the operational status (active or inactive) of each board and to its position in the chain (first, intermediate or last);
3. Write or read the boards via MCST/CBLT addressing.

An example of User procedures which can be used to perform a write access is:

`vme_write (address, data, addr_mode, data_mode),`

which contain the following parameters:

<i>Address:</i>	the complete address, i.e. Base Address + offset;
<i>Data:</i>	the data to be either written or read;
<i>Addr_mode:</i>	the addressing mode (A24 or A32);
<i>Data_mode:</i>	the data mode (D16, D32 or D64).




Fig. 4.4: MCST/CBLT Addressing Example

In the following two software examples using the above mentioned procedures are listed:

Example of Access via Base Address

```
vme_write (0xEE001004, 0xAA, A32, D16) /* set MCST Address=0xAA for board 1 */
vme_write (0xCC111004, 0xAA, A32, D16) /* set MCST Address=0xAA for board 2 */
vme_write (0xBC341004, 0xAA, A32, D16) /* set MCST Address=0xAA for board 3 */
```

```
vme_write (0xDD711004, 0xAA, A32, D16) /* set MCST Address=0xAA for board 4 */
vme_write (0xEE00101A, 0x02, A32, D16) /* set board 1 = First */
vme_write (0xCC11101A, 0x03, A32, D16) /* set board 2 = Active */
vme_write (0xBC34101A, 0x00, A32, D16) /* set board 3 = Inactive */
vme_write (0xDD71101A, 0x01, A32, D16) /* set board 4 = Last */
vme_write (0xAA001006, 0x80, A32, D16) /* set RESET MODE for all the boards */
```

4.2. Interrupter capability

The Mod. V1785 houses a RORAXtype VME INTERRUPTER. The INTERRUPTER responds to 8 bit, 16 bit and 32 bit interrupt acknowledge cycles by providing an 8-bit STATUS/ID on the VME data lines D00..D07.

4.2.1. Interrupt Status/ID

The interrupt STATUS/ID is 8-bit wide, and it is contained in the 8LSB of the Interrupt Vector Register (see § 4.12). The register is available at the VME address: Base Address + 0x100C.

4.2.2. Interrupt Level

The interrupt level corresponds to the value stored in the 3LSB of the Interrupt Level Register (see § 4.11). The register is available at the VME address: Base Address + 0x100A. If the 3LSB of this register are set to 0, the Interrupt generation is disabled.

4.2.3. Interrupt Generation

An Interrupt is generated when the number of events stored in the memory equals the value written in the Event Trigger Register at the VME address: Base Address + 0x1020 (see § 4.19). If the value in Event Trigger Register is set to 0 the interrupt is disabled (default setting).

4.2.4. Interrupt Request Release

The INTERRUPTER removes its Interrupt request when a Read Access is performed to the Output Buffer so that the number of events stored in the memory decreases and becomes less than the value written in the Event Trigger Register.

4.3. Data transfer capability

The internal registers are accessible in D16 mode, unless otherwise specified. Access in D32, BLT32, MBLT64, CBLT32 and CBLT64 is available for the data buffer.

4.4. Register address map

The Address map for the Model V1785 is listed in Table 4.2. All register addresses are referred to the Base Address of the board, i.e. the addresses reported in the Tables are the offsets to be added to the board Base Address.

The Table gives also information about the effects of RESET on the registers. In particular, column 2 through 4 refer to the following RESET operations:

- **D R → Data RESET;**
- **S R → Software RESET;**
- **H R → Hardware RESET.**

If a register has a mark in these columns, it means that the relevant RESET operation resets that register. For further details on the RESET Logic please refer to § 2.8.

Table 4.3 and Table 4.4 list register addresses (offset) in CBLT and MCST operations, respectively.

The ROM address map is reported in Table 4.5, p.55.

Table 4.2: Address Map for the Model V1785

Register content	DR	SR	HR	Address	Type	Access mode
Output Buffer	✓	✓	✓	0x0000÷0x07FF	Read only	D32/D64
Firmware Revision				0x1000	Read only	D16
Geo Address				0x1002	Read/Write	D16
MCST/CBLT Address			✓	0x1004	Read/Write	D16
Bit Set 1		✓ (*)	✓ (*)	0x1006	Read/Write	D16
Bit Clear 1		✓ (*)	✓ (*)	0x1008	Read/Write	D16
Interrupt Level		✓	✓	0x100A	Read/Write	D16
Interrupt Vector		✓	✓	0x100C	Read/Write	D16
Status Register 1		✓	✓	0x100E	Read only	D16
Control Register 1		✓ (*)	✓ (*)	0x1010	Read/Write	D16
ADER High			✓	0x1012	Read/Write	D16
ADER Low			✓	0x1014	Read/Write	D16
Single Shot Reset				0x1016	Write only	D16
MCST/CBLT Ctrl			✓	0x101A	Read/Write	D16
Event Trigger Register		✓	✓	0x1020	Read/Write	D16
Status Register 2		✓	✓	0x1022	Read only	D16
Event Counter_L	✓ (*)	✓ (*)	✓ (*)	0x1024	Read only	D16
Event Counter_H	✓ (*)	✓ (*)	✓ (*)	0x1026	Read only	D16
Increment Event				0x1028	Write only	D16
Increment Offset				0x102A	Write only	D16
Load Test Register				0x102C	Read/Write	D16
FCLR Window		✓	✓	0x102E	Read/Write	D16
Bit Set 2		✓	✓	0x1032	Read/Write	D16
Bit Clear 2		✓	✓	0x1034	Write only	D16
W Memory Test Address		✓	✓	0x1036	Write only	D16
Memory Test Word_High		✓	✓	0x1038	Write only	D16
Memory Test Word_Low				0x103A	Write only	D16
Crate Select		✓	✓	0x103C	Read/Write	D16
Test Event Write				0x103E	Write only	D16
Event Counter Reset				0x1040	Write only	D16
R Test Address		✓	✓	0x1064	Write only	D16
SW Comm				0x1068	Write only	D16
Slide Constant		✓	✓	0x106A	Read/Write	D16
AAD				0x1070	Read only	D16
BAD				0x1072	Read only	D16
Thresholds				0x1080 ÷ 0x10BF	Read/Write	D16

(*) not all bits are reset with the same type of RESET: see the description of the relevant register for details.

The ROM address map is from 0x8000 to 0xFFFF: refer to § 4.40.

Table 4.3: Address Map in CBLT operation

Register content	Address	Type	Access mode
Output Buffer	0x0000÷0x07FF	Read only	D32/D64

Table 4.4: Address Map in MCST operations

Register content	Address	Type	Access mode
Bit Set 1	0x1006	Write only	D16
Bit Clear 1	0x1008	Write only	D16
Interrupt Level	0x100A	Write only	D16
Interrupt Vector	0x100C	Write only	D16
Control Register 1	0x1010	Write only	D16
ADER High	0x1012	Write only	D16
ADER Low	0x1014	Write only	D16
Single Shot Reset	0x1016	Write only	D16
Event Trigger Register	0x1020	Write only	D16
Increment Event	0x1028	Write only	D16
Increment Offset	0x102A	Write only	D16
Load Test Register	0x102C	Write only	D16
Fast Clear Window	0x102E	Write only	D16
Bit Set 2	0x1032	Write only	D16
Bit Clear 2	0x1034	Write only	D16
W Memory Test Address	0x1036	Write only	D16
Memory Test Word_High	0x1038	Write only	D16
Memory Test Word_Low	0x103A	Write only	D16
Crate Select	0x103C	Write only	D16
Event Counter Reset	0x1040	Write only	D16
R Test Address	0x1064	Write only	D16
SW comm	0x1068	Write only	D16
Slide Constant	0x106A	Write only	D16
Thresholds	0x1080 ÷ 0x10BF	Write only	D16

4.5. Output Buffer Register

(Base Address + 0x0000 ÷ 0x07FC, read only)

This register allows the User to access the Multiple Event Buffer to readout the converted values.

The output buffer contains the output data organised in 32-bit words.

The data in the buffer are organised in events.

Each event consists of:

- the **header**, that contains the geographical address, the crate number and the number of converted channels;
- one or more **data words**, each of which contains the geographical address, the number of the channel, the Under-Threshold (UN) bit, the Overflow (OV) bit and the 12-bit converted value;
- the **End Of Block** (EOB), which contains the geographical address and the event counter.

31	30	29	28	27	26	25	24	23	22	21	20	19	18	17	16	15	14	13	12	11	10	9	8	7	6	5	4	3	2	1	0
GEO[4:0]	0	1	0		CRATE[7:0]		0	0		CNT[5:0]																					

Fig. 4.5: Output buffer: the Header

31	30	29	28	27	26	25	24	23	22	21	20	19	18	17	16	15	14	13	12	11	10	9	8	7	6	5	4	3	2	1	0	
GEO[4:0]	0	0	0		CHANNEL [2:0]	RG			UNOV																							

Fig. 4.6: Output buffer: the Data Word format

31	30	29	28	27	26	25	24	23	22	21	20	19	18	17	16	15	14	13	12	11	10	9	8	7	6	5	4	3	2	1	0	
GEO[4:0]	1	0	0																													

Fig. 4.7: Output buffer: the End Of Block

Header content:

The bits[31...27] contains the GEO address.

The bits[26..24] identify the type of word (010 → header);

The bits[23..16] identify the crate number according to the content of the Crate Select Register (see § 4.31).

The bits[13...8] contain the number of memorised channels.

Datum content:

The bits[31...27] contains the GEO address.

The bits[26..24] identify the type of word (000 → datum);

The bits[20..18] identify the channel which the data are coming from; the sequence is explained in § 4.5.1.

The bit [17] identifies the input RANGE of the channel (0 = High; 1 = Low)

The bit[13] is the UNDERTHRESHOLD bit:

- = 0 → the datum is over the threshold fixed in the relevant register (see § 4.35);
- = 1 → the datum is under the threshold fixed in the relevant register; it is actually possible to make the datum be written in the buffer even if it is under the threshold by using the bits 3 and 4 of the Bit Set 2 Register (see § 4.26);

The bit[12] is the OVERFLOW bit:

- = 0 → ADC not in overflow condition;
- = 1 → ADC in overflow;

The bits[11...0] contain the converted datum.

EOB content:

The bits[31...27] contains the GEO address.

The bits[26..24] identify the type of word (100 → EOB);

The bits[23..0] contain the 24-bit event counter value (see § 4.21).

The bits[31...27] always contains the GEO address (except for the not valid datum, see Fig. 4.8).

The bits[26..24] identify the type of word, according to the following:

- 010 → header;
- 000 → valid datum;
- 100 → end of block;
- 110 → not valid datum.
- *others* → reserved.

If a read access is performed to the buffer when it is empty, the readout will provide a NOT VALID DATUM arranged as shown in Fig. 4.8.

31	30	29	28	27	26	25	24	23	22	21	20	19	18	17	16	15	14	13	12	11	10	9	8	7	6	5	4	3	2	1	0

Fig. 4.8: Output buffer: not valid datum

4.5.1. Data storage

The sequence followed to store the data in the buffer is as follows:

HEADER

CHANNEL 0 High Range
CHANNEL 4 High Range
CHANNEL 0 Low Range
CHANNEL 4 Low Range
CHANNEL 1 High Range

.....

CHANNEL 3 Low Range
CHANNEL 7 Low Range
END OF BLOCK

Please note that some of the above channel data may be missing in the sequence: this is due either to overflow or under threshold conditions (which caused these data not to be stored), or to User's settings to kill some channels.

4.5.2. Data readout

A smart way to read out only the strictly necessary number of data is to set the Zero Suppression Threshold at \sim FSR/8⁵ (\sim 500 counts) on High Range Channels, which will be thus acquiring only pulses larger than 100 pC. Low Range Channels will be automatically suppressed by the Over Range function in the 100 pC \div 800 pC interval. Pulses larger than 800 pC will be rejected in any case (on both ranges).




Fig. 4.9: Data readout and suppression

⁵ It is suggested to set this threshold a little smaller than FSR/8, in order to avoid losing valid data

31 30 29 28 27 26 25 24 23 22 21 20 19 18 17 16 15 14 13 12 11 10 9 8 7 6 5 4 3 2 1 0																				
Write Pointer N GATE 5	GEO			0	1	0	CRATE NUMBER				0	0	MEM. CHANNELS (2)							
	GEO			0	0	0	CHANNEL (2 h.r.)					UN	OV	ADC COUNTS						
	GEO			0	0	0	CHANNEL (2 l.r.)					UN	OV	ADC COUNTS						
	GEO			1	0	0	EVENT COUNTER (m)													
Write Pointer N+1 GATE 7	GEO			0	1	0	CRATE NUMBER				0	0	MEM. CHANNELS (2)							
	GEO			0	0	0	CHANNEL (0 h.r.)					UN	OV	ADC COUNTS						
	GEO			0	0	0	CHANNEL (0 l.r.)					UN	OV	ADC COUNTS						
	GEO			1	0	0														

Fig. 4.10: Multi-Event Buffer: data structure example

N.B.: the GEO address must be written by the User via a write access to the relevant register (see § 4.6), since the module does not use the PAUX connector. If this operation is not performed, it will be not possible to identify which module the data are coming from when the CBLT access is used.

4.6. Firmware Revision Register

(Base Address + 0x1000, read only)

This register contains a 16-bit value identifying the firmware revision. The 16-bit value corresponds to 4 hexadecimal figures which give the firmware revision number. For example, in the figure is shown the register content for the firmware release:

Rev. 01.05

which presently is the latest one.

15	14	13	12	11	10	9	8	7	6	5	4	3	2	1	0	
0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	1	1
0x0				0x1				0x0				0x3				

Binary representation
Hexadecimal representation

Fig. 4.11: Firmware Revision Register

4.7. GEO Address Register

(Base Address + 0x1002, read/write)

This register contains the geographical address of the module. The register is filled up upon arrival of a RESET. The register content is the following:




Fig. 4.12: Geographical address register

GEO [4...0] corresponds to A23...A19 in the address space of the CR/CSR area: each slot has a relevant number whose binary encoding consists of the GEO ADDR 4 to 0.

The bits of the GEO Address register are set to 1 by default. In CBLT operation it is up to the User to write the correct GEO address of the module in this register before operating so that the GEO address will be contained in the HEADER and the EOB words for data identification.

N.B.: It is possible ONLY to perform a write access to the GEO Address Register for data identification during CBLT operation (see § 4.1.3), avoid to use the GEO Register for addressing purposes.

N.B.: after a write access to the GEO register, it is necessary to perform a reset to make the change active.

4.8. MCST/CBLT Address Register

(Base Address + 0x1004, read/write)

This register contains the most significant bits of the MCST/CBLT address of the module set via VME, i.e. the address used in MCST/CBLT operations. Refer to § 4.1.3 for details about MCST/CBLT addressing mode.

The register content is as follows:




Fig. 4.13: MCST/CBLT address register

Default setting (i.e. at power ON or after hardware reset) is 0xAA.

4.9. Bit Set 1 Register

(Base Address + 0x1006, read/write)

This register allows to set the RESET logic of the module and to enable the change of the base address via VME.

A write access with the bits to 1 sets the relevant bits to 1 in the register (i.e. writing 0x10 to this register sets the SEL ADDR bit to 1). A write access with the bits set to 0 does NOT clear the register content; in other words, when 1 is written into one particular bit, such bit is set to 1, if 0 is written, the bit remains unchanged. In order to clear the register content, the Bit Clear 1 Register must be used (see § 4.10).

A read access returns the status of this register.

The register content is the following:




Fig. 4.14: Bit Set 1 Register

BERR FLAG: Bus Error Flag Bit (meaningful in BLT/CBLT modes only). The User may set this flag for test purposes only. Its content is cleared both via an hardware and via a software reset.

= 0 board has not generated a Bus Error (default);
= 1 board has generated a Bus Error.

SELECT ADDRESS: Select Address bit.

= 0 base address is selected via Rotary Switch (default);
= 1 base address is selected via internal ADER registers.

SOFTW. RESET: Sets the module to a permanent RESET status. The RESET is released only via write access with the relevant bit set to 1 in the Bit Clear Register, see § 4.10.

This register is reset via a hardware reset (see § 2.8). Only the bit 3 (BERR FLAG) is reset both via hardware reset and software reset.

4.10. Bit Clear 1 Register

(Base Address + 0x1008, read/write)

This register allows to clear the bits in the above described Bit Set 1 Register. A write access with a bit set to 1 resets that bit, e.g. writing 0x8 to this register resets the BERR FLAG bit. A write access with the bits set to 0 does NOT clear the register content; in other words, when 1 is written into one particular bit, such bit is set to 0 (cleared), if 0 is written, the bit remains unchanged. The structure of the register is identical to the Bit Set 1 Register. A read access returns the status of the register.

4.11. Interrupt Level Register

(Base Address + 0x100A, read/write)

The 3 LSB of this register contain the value of the interrupt level (Bits 3 to 15 are meaningless). Default setting is 0x0. In this case interrupt generation is disabled.




Fig. 4.15: Interrupt Level Register

4.12. Interrupt Vector Register

(Base Address + 0x100C, read/write)

This register contains the STATUS/ID that the V1785 INTERRUPTER places on the VME data bus during the Interrupt Acknowledge cycle (Bits 8 to 15 are meaningless). Default setting is 0x00.




Fig. 4.16: Interrupt Vector Register

4.13. Status Register 1

(Base + 0x100E, read only)

This register contains information on the status of the module.

TERM ON and TERM OFF refer to the terminations of the CONTROL bus lines: the last module in a chain controlled via the front panel CONTROL connector must have these terminations ON, while all the others must have them OFF. The insertion or removal of the terminations is performed via internal DIP switches (see Fig. 3.2).

The BUSY and DATA READY signals are available both for the individually addressed module and as a global readout of a system of many units connected together via the CONTROL bus.




Fig. 4.17: Status Register

DREADY: Indicates that there are data (at least 1 event) in the Output Buffer.
 = 0 No Data Ready;
 = 1 Data Ready.

GLOBAL DREADY: Indicates that at least one module in the chain has data in the Output Buffer (OR of the READY+ signal of each module in the chain).

- = 0 No module has Data Ready;
- = 1 At least one module has Data Ready.

BUSY: Busy status indicates that either a conversion is in progress or the board is resetting or the Output Buffer is full or the board is in MEMORY TEST mode.

- = 0 Module not Busy;
- = 1 Module Busy.

GLOBAL BUSY: Indicates that at least a module in a chain is BUSY (OR of the BUSY+ signal of each module in the chain).

- = 0 No module is Busy;
- = 1 At least one module is Busy.

AMNESIA: not used

PURGED: during a CBLT operation it indicates that the board is purged, i.e. the board has finished to send data.

- = 0 the board is not purged;
- = 1 the board is purged.

TERM ON: Termination ON bit.

- = 0 not all Control Bus Terminations are ON.
- = 1 all Control Bus Terminations are ON.

TERM OFF: Termination OFF bit.

- = 0 not all Control Bus Terminations are OFF
- = 1 all Control Bus Terminations are OFF.

EVRDY: is a flag for the Event Trigger Register.

- = 0 (default) indicates that the number in the Event Trigger Register (see § 4.19) is smaller than the number of events stored in the memory;
- = 1 indicates that the number in the Event Trigger Register (see § 4.19) is greater than or equal to the number of events stored in the memory and an interrupt request has been generated with interrupt level different from 0 (see § 4.2.3).

N.B.: the condition in which both TERM ON and TERM OFF bits are equal to 0 indicates an uncommon termination status, e.g. some terminations are on and other are off.

4.14. Control Register 1

(Base Address + 01010, read/write)

This register allows performing some module's general settings.




Fig. 4.18: Control Register 1

BLKEND:

End of Block bit. Used in Block Transfer modes only.

= 0 The module sends all requested data to the CPU; when the Output Buffer is empty it will send no valid data. If BERR_VME is enabled (see bit 5 below, BERR ENABLE), a Bus Error is generated with the readout of the last word in the Output Buffer (default).

= 1 The module sends all data to the CPU until the first EOB word (end of first event) is reached; afterwards it will send no valid data. If BERR_VME is enabled, a Bus Error is generated at the readout of the EOB word.

PROG RESET:

Programmable Reset Mode setting bit.

= 0 the front panel RESET acts only on data (data reset, default);
= 1 the front panel RESET acts on the module (software reset).

N.B. This bit is cleared only via hardware reset.

BERR ENABLE:

Bus Error enable bit. Used in Block Transfer mode only.

= 0 the module sends a DTACK signal until the CPU inquires the module (default);

= 1 the module is enabled to generate a Bus error to finish a block transfer.

ALIGN 64:

Allows to add a 32 bit dummy-word (marked as not valid datum, see § 4.5) to an event which is made up of an odd number of words during BLT32 and CBLT32 data readout. In fact some 64 bit CPU's cut off the last 32 bit word of a transferred block if the number of words composing such block is odd, so it is necessary to add a dummy word (which will be then eventually removed via software) in order to avoid data loss. It is used in BLT32 and CBLT32 and is available in Firmware Rev. 6.02 and later.

= 0 no dummy word added (default);

= 1 dummy word added when the number of words is odd.

(Bits 7 to 15 are meaningless).

This register is reset both via software and via hardware reset (see § 2.8), except for the bit 4 (PROG RESET) which is reset only via hardware reset.

4.15. Address Decoder High Register

(Base Address + 0x1012, read/write)

This register contains the A31...A24 bits of the module's address: it can be set via VME for a relocation of the module's Base Address. The register content is as follows:




Fig. 4.19: ADER HIGH Register

4.16. Address Decoder Low Register

(Base Address + 0x1014 read/write)

This register contains the A23...A16 bits of the module's address: it can be set via VME for a relocation of the module's Base Address. The register content is as follows:




Fig. 4.20: ADER LOW Register

4.17. Single Shot Reset Register

(Base Address + 0x1016, write only)

A write access to this dummy register performs a module reset. This register must be used very carefully and for debugging purposes only. In order to reset the board, it is recommended to use the Bit Set 1 Register (see § 4.9).

4.18. MCST/CBLT Control Register

(Base Address + 0x101A, write only)

This register allows performing some general MCST/CBLT module's settings.




Fig. 4.21: MCST Address Register

LAST_BOARD	Last Board flag bit (valid in CBLT and MCST modes only)
FIRST_BOARD	First Board flag bit (valid in CBLT and MCST modes only)

The status of the boards according to the bit value is the following:

BOARD STATUS	FIRST BOARD bit	LAST BOARD bit
Board disabled in CBLT or MCST chain	0	0
First board in CBLT or MCST chain	1	0
Last board in CBLT or MCST chain	0	1
Active intermediate board in CBLT or MCST chain	1	1

(Bits 2 to 15 are meaningless).

4.19. Event Trigger Register

(Base Address + 0x1020, read/write)

This register contains a 5-bit value set by the User: when the number of events stored in the memory equals this value an interrupt request is generated.

Default setting is 0: in this case the interrupt generation is disabled. See also § 4.2.




Fig. 4.22: Event Trigger Register

4.20. Status Register 2

(Base Address + 0x1022, read only)

This register contains further information on the status of the module's output buffer and on the type of piggy back plugged into the main board.




Fig. 4.23: Status Register 2

BUFFER EMPTY:

Indicates if the output buffer is empty.

=0 buffer not empty;
=1 buffer empty.

BUFFER FULL:

Indicates if the output buffer is full.

=0 buffer not full;
=1 buffer full.

CSEL1, CSEL0, DSEL1, DSEL0: Indicate the type of piggy-back plugged into the board.

4.21. Event Counter_Low Register

(Base Address + 0x1024, read only)

It contains the event counter 16 LSBs. The event counter can work in two different ways (see also § 2.6):

1. it counts all events;
2. it counts only the accepted events.

The two modes can be selected via the Bit 14 (ALL TRG) of the Bit Set 2 Register (see § 4.26).

EVENT CNT LOW:16 LSB of the 24-bit Event Counter.




Fig. 4.24: Event Counter Low Register

This register is reset via the Event Counter Reset Register (see § 4.33) or via a software or hardware reset (see § 2.8). However, if the event counter is set so as to work as relative counter (i.e. it counts only the accepted events), this register is reset also with a data reset (see § 2.8).

4.22. Event Counter_High Register

(Base + 0x1026, read only)

It contains the 8 MSB of the 24-bit event counter. The event counter can work in two different ways (see also § 2.6):

1. it counts all events;
2. it counts only the accepted events.

The two modes can be selected via the Bit 14 (ALL TRG) of the Bit Set 2 Register (see § 4.26).

EVENT CNT HIGH: 8 MSB of the 24-bit Event Counter.




Fig. 4.25: Event Counter High Register

This register is reset via the Event Counter Reset Register (see § 4.33) or via a software or hardware reset (see § 2.8). However, if the event counter is set so as to work as relative counter (i.e. it counts only the accepted events), this register is reset also with a data reset (see § 2.8).

4.23. Increment Event Register

(Base Address + 0x1028, write only)

A write access to this dummy register sets the readout pointer on the next event in the output buffer (at the first address).

In particular, if the bit 11 (AUTO INCR) of the Bit Set 2 Register is set to 0 (see § 4.26), the readout pointer is no more automatically incremented, but it can be incremented via a write access to this register or to the Increment Offset Register (see below).

4.24. Increment Offset Register

(Base Address + 0x102A, write only)

A write access to this dummy register increments the readout pointer of one position (next word, same event if EOB is not encountered; next event if EOB is encountered). In particular, if the bit 11 (AUTO INCR) of the Bit Set 2 Register is set to 0 (see § 4.26), the readout pointer is no more automatically incremented, but it can be incremented via a write access to this register or to the Increment Event Register (see above).

4.25. Fast Clear Window Register

(Base Address + 0x102E, read/write)

For the definition of the Fast Clear window refer to Fig. 2.8. By writing a 10 bit number N to this register, it is possible to set the Fast Clear window width T_{FC} in the range 7÷38.5 μ s (1/32 μ s steps) according to the following relation:

$$T_{FC} (\mu\text{s}) = N \times T_{CLOCK} + 7 \mu\text{s}$$

where $T_{CLOCK} = 1/32 \mu\text{s}$.

15	14	13	12	11	10	9	8	7	6	5	4	3	2	1	0
N (number of 32 MHz clock cycles)															

Fig. 4.26: Fast Clear Window Register

Please note that the maximum allowed value for N is 3F0 which leads to $T_{FC} = 38.5 \mu\text{s}$

4.26. Bit Set 2 Register

(Base Address + 0x1032, read/write)

This register allows to set the operation mode of the module. A write access with a bit to 1 sets the relevant bit to 1 in the register. A write access with the bit set to 0 does not clear the register content, the Bit Clear 2 Register must be used (see § 4.27). A read access returns the status of the register. The register content is the following:




Fig. 4.27: Bit set 2 register

MEM TEST: Test bit: allows to select the Random Memory Access Test Mode (see § 5.7.1).
 =0 normal mode (default);
 =1 Random Memory Access Test Mode selected: it is possible to write directly into the memory.

OFFLINE: Offline bit: allows to select the ADC controller's status.
 =0 ADC controller online (default);
 =1 ADC controller offline: no conversion is performed.

CLEAR DATA: Allows to generate a reset signal which clears the data, the write and read pointers, the event counter and the peak sections.
 =0 no data reset is generated (default);
 =1 a data reset signal is generated.

OVER RANGE: Allows to disable overflow suppression (see also § 2.3).

- =0 over range check enabled: only the data not causing the ADC overflow are written into the output buffer (overflow suppression; default);

=1 over range check disabled: all the data are written into the output buffer (no overflow suppression).

- LOW THRESHOLD:** Allows to disable zero suppression (see also § 2.3).
 - =0 low threshold check enabled: only data above the threshold are written into the output buffer (zero suppression; default);
 - =1 low threshold check disabled: all the data are written into the output buffer (no zero suppression).

- TEST ACQ:** Allows to select the Acquisition Test Mode (see § 5.7.2).
 - =0 normal operation mode, i.e. the data to be stored in the buffer are the real data (default);
 - =1 Acquisition Test Mode selected, i.e. the data to be stored in the buffer are taken from an internal FIFO (Test Event Write Register, see § 4.32).

- SLIDE ENABLE:** Allows to enable/disable the sliding scale.
 - =0 the sliding scale is disabled and the DAC of the sliding scale is set with a constant value (Slide Constant, see § 4.36);
 - =1 the sliding scale is enabled (default).

- STEP TH:** Allows to set the zero suppression threshold resolution (firmware release 5.1 and later, see § 2.3 for details)
 - =0 ADC CONVERTED VALUE < THRESHOLD VALUE x 16
 - =1 ADC CONVERTED VALUE < THRESHOLD VALUE x 2

- AUTO INCR:** Allows to enable/disable the automatic increment of the readout pointer.
 - =0 the read pointer is not incremented automatically but only by a write access to the Increment Event or Increment Offset Registers (see § 4.23 and 4.24);
 - =1 the read pointer is incremented automatically (default).

- EMPTY PROG:** Allows to choose if writing the header and EOB when there are no accepted channels.
 - =0 when there are no accepted channels, nothing is written in the output buffer (default).
 - =1 when there are no accepted channels, the Header and the EOB are anyway written in the output buffer.

- SLIDE_SUB ENABLE:** Allows to change operation mode for the sliding scale.
 - =0 the sliding scale works normally (default);
 - =1 the subtraction section of the sliding scale is disabled (test purposes only).

- ALL TRG:** Allows to choose how to increment the event counter.
 - =0 event counter incremented only on accepted triggers.
 - =1 event counter incremented on all triggers (default).

4.27. Bit Clear 2 Register

(Base Address + 0x1034 write only)

This register allows clearing the bits of the Bit Set 2 Register (§ 4.26). A write access with a bit set to 1 resets that bit, e.g. writing 0x4 to this register resets the CLEAR DATA bit. A write access with the bits set to 0 does NOT clear the register content. The structure of the register is identical to the Bit Set 2 Register.

4.28. W Memory Test Address Register

(Base Address + 0x1036 write only)

This register contains the memory address on which data can be written for the memory test.




Fig. 4.28: W Memory Test Address Register

N.B.: The output buffer is a FIFO, so the read address (R Test Address Register) must be different from the write address (W Test Address Register).

4.29. Memory Test Word_High Register

(Base Address + 0x1038 write only)

The Memory Test Word is a 32-bit word used for the memory test. The higher 16 bits are set via this register, while the lower 16 bits are set via the Test Word_Low Register.

These registers are used in TEST mode as follows:

1. set the module in test mode (see bit 0 of the Bit Set 2 Register, § 4.26);
2. write the memory address (see § 4.26),
3. write the 16 MSBs in the TESTWORD_HIGH register;
4. write the 16 LSBs in the TESTWORD_LOW register;

With the latter operation, the 32-bit pattern is transferred to the memory. If operations 3. and 4. are inverted, the content of the 16 MSBs may be meaningless.




Fig. 4.29: Test Word_High Register

4.30. Memory Test Word_Low Register

(Base Address + 0x103A write only)

This register allows to set the lower 16 bits of the Test Word (see above).




Fig. 4.30: Test Word_Low Register

4.31. Crate Select Register

(Base Address + 0x103C read/write)

This register contains the crate number which the board is plugged into. This register must be filled at board initialisation and will be part of the data word (see § 4.5).




Fig. 4.31: Crate Select Register

4.32. Test Event Write Register

(Base Address + 0x103E write only)

This register is used in Acquisition Test Mode and its content constitutes the test event to be written in the output buffer.

A write access to this register allows the User to write a set of 16 data into a 16-word FIFO. As the Bit 6 (TEST ACQ) of the Bit Set 2 Register (see § 4.26) is set to 1 and the Acquisition Test Mode is consequently selected, these data are directly written in the output buffer constituting an event which can be used to test the module and/or the acquisition software.

Each 16-bit test word (see the figure below) contains a 12-bit value, acting as the ADC converted value, and an OV bit which indicates the possible overflow.

The 16 test data, corresponding to the data from the 8 channels, must be written in this FIFO in the same order as they are read from the output buffer, that is:

- test datum for the CHANNEL 0 High Range
- test datum for the CHANNEL 4 High Range
- test datum for the CHANNEL 0 Low Range
-
- test datum for the CHANNEL 3 Low Range
- test datum for the CHANNEL 7 Low Range

For further details on the use of this register in Acquisition Test Mode please refer to § 5.7.2.

N.B.: please note that the User must write at least and not more than 16 test words. Actually, since the words are written in a FIFO, if the User writes less than 16 words, some words will be not defined; on the other hand, if the User writes more than 16 words, some words will be overwritten.




Fig. 4.32: Test Event Write Register

4.33. Event Counter Reset Register

(Base Address + 0x1040 write only)

A VME write access to this dummy register clears the Event Counter.

4.34. R Memory Test Address Register

(Base Address + 0x1064 write only)

This register contains the output buffer address from which data can be read for the memory test.




Fig. 4.33: R Memory Test Address Register

N.B.: The output buffer is a FIFO, so the read address (R Test Address Register) must be different from the write address (W Test Address Register).

4.35. SW Comm Register

(Base Address + 0x1068 write only)

A write access to this dummy register causes a conversion for test purposes.

4.36. Slide constant Register

(Base Address + 0x106A read/write)




Fig. 4.34: Slide Constant Register

This register contains a 8-bit value corresponding to the constant to which is set the sliding scale DAC when the sliding scale is disabled by means of the SLD_ENABLE bit of the Bit Set 2 Register (refer to § 4.26).

4.37. AAD Register

(Base Address + 0x1070 read only)

This register contains the value converted by the ADC of the Block A (refer to the block diagram of Fig. 1.1).




Fig. 4.35: AAD Register

4.38. BAD Register

(Base Address + 0x1072 read only)

This register contains the value converted by the ADC of the Block B (Refer to the block diagram of Fig. 1.1).




Fig. 4.36: BAD Register

4.39. Thresholds Memory

(Base Address + 0x1080 ÷ 0x10BC read/write)

This register contains the low threshold and kill option for each channel. The address is different for each channel (ch0high → 0x1080, ch0low → 0x1084, ..., ch7high → 0x10BA, ch7low → 0x10BC).

Each threshold is as shown in the figure:




Fig. 4.37: Threshold Register

KILL (K): allows to abort memorisation of the data from the relevant channel.

= 0 channel data are memorised;

= 1 channel data memorisation is aborted.

THRESHOLD VALUE: this is a 8-bit value which is compared with the 8MSB of the 12-bit value to be memorised.

Default settings are not defined.

Please note that the KILL option can be used to disable some channels.

N.B.: the threshold values are reset only with a hardware reset and when the board is switched off.

4.40. ROM memory

(Base Address + 0x8000 ÷ 0xFFFF, read only)

It contains some useful information according to the table below, such as:

- OUI:** manufacturer identifier (IEEE OUI);
- Version:** purchased version of the Mod.V1785;
- Board ID:** Board identifier (1785);
- Revision:** hardware revision identifier;
- Serial MSB:** serial number (MSB);
- Serial LSB:** serial number (LSB).

Table 4.5: ROM Address Map for the Model V1785

Description	Address	Content (*)
OUI MSB	0x8026	0x00
OUI	0x802A	0x40
OUI LSB	0x802E	0xE6
Version	0x8032	0x00
BOARD ID MSB	0x8036	0x00
BOARD ID	0x803A	0x06
BOARD ID LSB	0x803E	0xF9
Revision	0x804E	0x00
Serial MSB	0x8F02	0x00
Serial LSB	0x8F06	0x02

(*) the example of content for the relevant register refers to the Mod.V1785 (serial number: 2; hardware revision: 0).

5. Operating modes

5.1. Safety information

This section contains the fundamental safety rules for the installation and operation of the module.

Read thoroughly this section before starting any procedure of installation or operation of the product.

5.2. Product Damage Precautions

Use Proper Power Source.

Do not operate this product from a power source that applies more than the voltage specified.

Provide Proper Ventilation.

To prevent product overheating, provide proper ventilation.

Do Not Operate With Suspected Failures.

If you suspect there is damage to this product, have it inspected by qualified service personnel.

5.3. Installation

The V1785 board can be inserted into both standard and V430 VME 6U crates. Please note that the board supports live insertion/extraction into/from the crate, i.e. it is possible to insert or extract it from the crate without turning the crate off. Moreover, it is possible to switch the board off by the relevant PWR switch (see § 3.5.2) without cutting the interrupt chain off.

**CAUTION**

ALL CABLES MUST BE REMOVED FROM THE FRONT PANEL
BEFORE EXTRACTING THE BOARD FROM THE CRATE!

5.4. Power ON sequence

To power ON the board follow this procedure:

1. insert the V1785 board into the crate: as the board is inserted, the OVC PWR green LED lights up indicating that the board is powered;
2. after a short time the BUSY and DRDY LEDs will light off: this indicates that the board is ready to acquire data.

N.B.: if the OVC PWR LED becomes orange instead of being green, there is an overload and the over-current protection is now running. In order to acquire data, it is necessary to remove the overload source, then turn the board off and switch it on again. Sometimes, it may happen that the OVC PWR LED is orange as soon as the board is inserted in the crate: this is due to the fact that the board has been just misplaced into the crate. In this case, extract the board and insert it again into the crate.

5.5. Power ON status

At power ON the module is in the following status:

- the Event Counter is set to 0;
- the Output buffer is cleared;
- the Read and Write Pointer are cleared (i.e. Buffer 0 is pointed);
- the Interrupt Level is set to 0x0 (in this case interrupt generation is disabled) and the Interrupt Vector is set to 0x0;
- the values in the threshold memory are not defined (see § 4.39);
- the MCST/CBLT address is set to 0xAA.

Moreover, all other registers marked in the column HR (Hardware RESET) in Table 4.2 are cleared or set to the default value.

At power on or after a hardware reset (see § 2.8) the module must thus be initialised.

5.6. Operation sequence

After the power ON sequence the module is in the status described above.

Please note that the threshold values are not defined after power ON and consequently before starting the operation of the module it is necessary to set a threshold value for each channel in the Threshold memory (refer to § 4.39).

If the module is not BUSY, a GATE input pulse causes the following:

1. starts the Charge to Amplitude Conversion;
2. increments the event counter according to the User's settings (see § 2.6);
3. sets the BUSY output signal to 1.

If neither RESET nor FAST CLEAR occur (refer to § 2.8 and § 2.9) to abort the Charge to Amplitude Conversion, the control logic starts the following conversion sequence:

1. The outputs of the PEAK sections are multiplexed and sampled;
2. The control logic checks if there are accepted data among the converted values, according to the User's settings (zero suppression, overflow suppression and KILL option: see § 2.3 and § 2.3):
 - a) if there are accepted data, these are stored in the active event buffer together with a Header and an EOB;
 - b) if there are no accepted data and the EMPTY ENABLE bit of the Bit Set 2 Register is set to 0 (default setting, see § 4.26), no data will be written in the output buffer.
 - c) if there are no accepted data and the EMPTY ENABLE bit of the Bit Set 2 Register is set to 1 (see § 4.26), the Header and EOB only will be written in the output buffer.
3. If the MEB is not full the PEAK sections and the BUSY are cleared and the module is ready for the next acquisition; if the MEB is full the module doesn't accept any GATE and BUSY is not cleared.

5.7. Test Modes

Two different test modes can be enabled:

- **Random Memory Access Test Mode**
- **Acquisition Test Mode**

The first test mode operation is enabled via the Bit 0 of the Bit Set 2 Register and allows to write directly into the buffer.

The second test mode is enabled via the Bit 6 of the Bit Set 2 Register and allows to test the whole acquisition system by writing a set of 32 data in an internal FIFO which are then transferred to the output buffer at each GATE pulse for the readout.

The test modes will be described in detail in the following subsections.

5.7.1. Random Memory Access Test Mode

This test mode allows the User to write and read a word in the output buffer.

To perform such test follow these steps:

1. Set to 1 the Bit 0 of the Bit Set 2 Register (see § 4.26);
2. Write into the W Memory Test Address Register (see § 4.26) the 11-bit address where to write the test word;
3. Write the high and low part of the 32-bit test word respectively in the Testword_High and Testword_Low Registers (see § 4.29 and § 4.30). As the Testword_Low register is accessed, the whole test word is written into the memory;
4. Write in the R Test Address Register (see § 4.34) the 11-bit reading memory address and read out the buffer; please note that this address must be different from the write address written in the W Memory Test Address Register.

N.B.: please note that the R Memory Test Address must be different from the W Memory Test Address at any step of the procedure. If the User tries to write an

address in one of these registers that is equal to the address contained in the other register, write cycles (step 3. above) will not write the correct value.

5.7.2. Acquisition Test Mode

This test mode allows the User to simulate the real operation of the board without using any channel input signals but just writing the data into a FIFO via an appropriate register (Test Event Write Register, see § 4.32) and reading them after a GATE signal.

To operate the acquisition test follow these steps:

1. Set to 1 the Bit 6 (TEST ACQ) of the Bit Set 2 Register (see § 4.26); this action selects the Acquisition Test Mode and resets the write pointer in the FIFO;
2. Set to 0 the Bit 6 (TEST ACQ) of the Bit Set 2 Register (see § 4.26); this action resets the read pointer in the FIFO and releases the write pointer;
3. Write 16 data words (each word consisting of a 13-bit word, corresponding to the ADC converted value, + the overflow bit, see § 4.32) in the Test Event Write Register (Base Address + 0x103E). These 16 data constitute the event to obtain as output of the 8 channels (dual range). The 16 test data must be written in this FIFO in the same order as they will be read from the output buffer, that is:
 - test datum for the channel 0 high range
 - test datum for the channel 4 high range
 - test datum for the channel 0 low range
 -
 - test datum for the channel 7 high range
 - test datum for the channel 3 low range
 - test datum for the channel 7 low range

N.B.: please note that the User must write at least and not more than 16 test words. Actually, since the words are written in a circular FIFO, if the User writes less than 16 words, some words will be not defined; on the other hand, if the User writes more than 16 words, some words will be overwritten.

4. Set to 1 the Bit 6 (TEST ACQ) of the Bit Set 2 Register (see § 4.26); this action resets again the write pointer in the FIFO and releases the read pointer;
5. Send a set of GATE input signals: at each GATE signal the data previously written in the FIFO will be transferred to the output buffer. The data will be read via VME in the same order as they were written into the FIFO:
 - test datum for the channel 0 high range
 - test datum for the channel 4 high range
 - test datum for the channel 0 low range
 -
 - test datum for the channel 7 high range
 - test datum for the channel 3 low range
 - test datum for the channel 7 low range

N.B.: To operate in normal mode again, the Bit 6 of the Bit Set 2 Register must be set again to 0.

5.8. Block Transfer Mode

The module supports the Standard BLT32 and MBLT64 modes.

A standard readout in Block Transfer mode, for example, consists of a readout of the Header for the relevant event and a Block Transfer readout of the number of data words relative to the event (the number of data words referring to the event is the CNT number in the Header, see § 4.5).

A more efficient readout in Block Transfer mode can be performed by using the BLOCK END and BERR-ENABLE bits of the Control Register 1 (see § 4.14).

Some examples of this type of readout in Block Transfer mode are as follows:

Example A: BLOCK END = 0, BERR_ENABLE = 0;
A Block Transfer readout of 32x34 words (32 events max., each event 34 words max.) allows the readout of all data stored in the buffer: as the buffer is empty, the module will send only not valid data.

Example B: BLOCK END = 0, BERR_ENABLE = 1;
A Block Transfer readout of 32x34 words (32 events max., each event 34 words max.) allows the readout of all events stored in the buffer: as the buffer is empty, a BERR is generated.

Example C: BLOCK END = 1, BERR_ENABLE = 0;
A Block Transfer readout of 34 words (each event 34 words max.) allows the readout of one complete event: after the readout of the EOB the module will send only not valid data.

Example D: BLOCK END = 1, BERR_ENABLE = 1;
A Block Transfer readout of 34 words (each event 34 words max.) allows the readout of one complete event: as the EOB is encountered, a BERR is generated.

The use of the BERR_ENABLE bit (Examples B and D above) is suggested only if the VME CPU can handle the Bus Error (BERR) in an effective way.

N.B.: Please note that, according to the VME standard, a Block Transfer readout can be performed with 256 read cycles maximum: as a consequence, a readout with a greater number of read cycles may require more BLT operations.

This limit is not due to the board itself but only to the VME standard: if it is possible to disable or delay the timeout of the BUS Timer (BTO(x)), a Block Transfer readout with more than 256 read cycles can be performed as well.

5.9. Advanced Setting and Readout Modes

Chained Block Transfer (CBLT) and Multicast (MCST) operations allow to enhance the set and readout time of the 8 channels. These operations allow accessing several boards at the same time: CBLT operations are used for reading cycles only, while MCST operations are used for write cycles only. For further details on the CBLT/MCST addressing mode please refer to § 4.1.3 and § 4.1.4.

In order to perform CBLT and MCST operations, the higher Base Address bits of all the involved modules (i.e. bits 31 to 24) must be set in common to all boards via the

MCST/CBLT Address Register (see § 4.8). This means that all boards must have the same setting on bits 31 to 24.

The resulting MCST (CBLT) Base Address for all boards is:

MCST (CBLT) Base Address = %NN000000,

Once the addresses have been set, the first and last board in a chain must have, respectively, only the FIRST_BOARD and only the LAST_BOARD bit set to 1 in the MCST Control Register (see § 4.8). Conversely, all intermediate boards must have both the FIRST_BOARD and the LAST_BOARD bits set either to 1 or to 0.

5.9.1. Chained Block Transfer Mode

Once set the address of the boards as described in the above section, the boards can be accessed in Chained Block Transfer mode (CBLT, see [5]). This mode allows for sequential readout of a certain number of contiguous boards in a VME crate. A CBLT access is allowed with the BLT32 and MBLT64 address modifiers only (CBLT32 and CBLT64 accesses respectively).

N.B.: The CBLT operation can be performed only for the readout of the Multi-Event Buffer: its address in CBLT mode corresponds to the set of offsets listed in Table 4.3 to be added to the address, common to all boards, set by the User via the MCST/CBLT Address Register which contains the most significant bits of the address (see § 4.8).

The User must perform a number of CBLT accesses that allows for the readout of all data in all boards of the chain in all possible occupancy conditions. E.g.: if the User has a chain of 10 boards, the total number of words for a given event lies between 0 (i.e. no data) and $34 \times 10 = 340$ 32-bit words (i.e. each board has an event, each event consists of a Header + 32 data + End of Block). In order to be sure that a BERR is generated, the User must thus perform 11 CBLT accesses of 34-word each.

In CBLT32 mode the first board of the chain starts sending data (if there are any, i.e. if it is not purged, see § 4.13); as it has sent all data and the EOB is met, the board becomes purged, i.e. the relevant bit (PURGED) of the Status Register 1 is set to 1. This implies that the board will not be involved in the CBLT access any more since it has already sent all the required data. At this point the IACKOUT line is asserted and the next board, if not purged, starts sending data. As the last board receives the token and is purged, it asserts a BERR which acts as a data readout completion flag.

In CBLT64 mode the accesses work as in the CBLT32 one, except for the fact that the address is acknowledged during the first cycle and consequently a DTACK is asserted at least once.

In CBLT mode the Read Pointer must be incremented automatically: if the AUTOINC_ENABLE bit is set to 1 in the Bit Set 2 Register (see § 4.26), the Read Pointer is automatically incremented with the readout of the End Of Block word of each board; if the AUTOINC_ENABLE bit is set to 0, the Read Pointer is not automatically incremented and only the Header of the first word is read.

N.B.: Please note that, according to the VME standard, a Chained Block Transfer readout can be performed with 256 read cycles maximum: as a consequence, a readout with a greater number of read cycles may require more CBLT operations.

This limit is not due to the board itself but only to the VME standard: it is actually possible to perform a CBLT readout with more than 256 read cycles if the timeout of the BUS Timer (BTO(x)) is disabled or delayed.

If the latter action is not allowed and the CBLT readout stops before having read all data, the new CBLT cycle will start from where the token was left in the previous cycle: this goes on until the last board is reached and all data read, so that a BERR is generated.

5.9.2. Multicast Commands

Once set the address of the boards as described in § 5.9, the boards can be accessed in Multicast Commands (MCST) mode. The MCST mode allows to write in the registers of several boards at the same time by accessing a dummy Address only once. The latter is composed by the MCST Base Address plus the offset of the relevant register, according to the list shown in Table 4.4. Refer to § 4.1.3 for details on MCST addressing mode.

MCST access can be meaningless (even if possible) for the setting parameters depending on the individual channel characteristics.

N.B.: the MCST/CBLT Address Register must NEVER be accessed in MCST mode since this can affect the CBLT and MCST operations themselves.

6. References

- [1] C. Cottini, E. Gatti, V. Svelto, "A new method of analog to digital conversion", NIM vol. 24 p.241, 1963.
- [2] C. Cottini, E. Gatti, V. Svelto, "A sliding scale analog to digital converter for pulse height analysis", in Proc. Int. Symp. Nuclear, Paris, Nov. 1963.
- [3] G. Bianchetti et al., "Specification for VMEbus CRATE Type V430", CERN-EP, January 1990.
- [4] VME64 extensions draft standard, Vita 1.1-199x, draft 1.8, June 13,1997.
- [5] VMEBus for Physics Application, Recommendations & Guidelines, Vita23-199x, draft 1.0, 22 May 1997.

Both documents are available from URL: <http://www.vita.com>

Document type:
User's Manual (MUT)

Title:
Mod. V1785 8 Ch. Dual Range Peak ADC

Revision date:
26/07/2011

Revision:
5

APPENDIX A

VME interface timing

A.1 VME Cycle timing in D16/D32 mode

The figure below reports the Data Select (DS0 or DS1) - Data Acknowledge (DTACK) VME cycle in D16 mode and relative timing.

The theoretical minimum duration of the VME cycle in D16/D32 mode is 120+60 ns.




Fig.A.1: VME cycle timing in D16 mode

A.2 VME Cycle timing in BLT / CBLT mode

The figure below reports the Data Select (DS0 or DS1) - Data Acknowledge (DTACK) VME cycle in BLT / CBLT mode and relative timing.

The theoretical minimum duration of the VME cycle in BLT/CBLT mode is 60+15 ns.




Fig.A.2: VME cycle timing in BLT/CBLT mode

A.3 VME Cycle timing in MBLT / CBLT64 mode

The figure below reports the Data Select (DS) - Data Acknowledge (DTACK) VME cycle in MBLT / CBLT64 mode and relative timing.

The theoretical minimum duration of the VME cycle in MBLT/CBLT64 mode is 120+15 ns.




Fig.A.3: VME cycle timing in MBLT/CBLT64 mode

Document type:
User's Manual (MUT)

Title:
Mod. V1785 8 Ch. Dual Range Peak ADC

Revision date:
26/07/2011

Revision:
5

**CAEN S.p.A.**

Via Vetraia 11
55049 - Viareggio
Italy
Phone +39 0584 388 398
Fax +39 0584 388 959
info@caen.it
www.caen.it

**CAEN GmbH**

Brunnenweg 9
64331 Weiterstadt
Germany
Tel. +49 (0)212 254 4077
Mobile +49 (0)151 16 548 484
info@caen-de.com
www.caen-de.com

CAEN Technologies, Inc.

1 Edgewater Street - Suite 101
Staten Island, NY 10305
USA
Phone: +1 (718) 981-0401
Fax: +1 (718) 556-9185
info@caentechnologies.com
www.caentechnologies.com

CAENspa INDIA Private Limited

B205, BLDG42, B Wing,
Azad Nagar Sangam CHS,
Mhada Layout, Azad Nagar, Andheri (W)
Mumbai, Mumbai City,
Maharashtra, India, 400053
info@caen-india.in
www.caen-india.in



Copyright © CAEN SpA. All rights reserved. Information in this publication supersedes all earlier versions. Specifications subject to change without notice.